

80-003387 DB

RECEIVED

MAY 03 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 13:22

Date of Inspection: 05/03/2017

Serial Number: 80-003387
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) <i>201604C</i> Lot#:201611B Exp: 11/15/2018 <i>04/05/2018</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.050	0.079	0.198	0.079


Number of Simulators Used: 5

Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

ROGER G SKIPPER

05/03/2017
Date

*Reviewed
5-3-17
M/STP. T.A. B...
4/5*