

80-003945

RECEIVED

MAR 07 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 10:53

Date of Inspection: 03/07/2017

Serial Number: 80-003945
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22315080A2 Exp: 09/05/2017
0.000	0.050	0.079	0.196	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.080	0.198	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L Franz

GORDON L FRANZ

Signature and Printed Name

03/07/2017
Date

80-003945 *200*

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
FEB 28 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 10:48

Date of Inspection: 02/27/2017

Serial Number: 80-003945
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22315080A2 Exp: 09/05/2017
0.000	0.050	0.080	0.197 / 0.200	0.078
0.000	0.050	0.081	0.199 / 0.200	0.078
0.000	0.050	0.080	* / 0.201	0.078

Number of Simulators Used: 5

Remarks:

20: RFI Detect.

* OFFICER CAME INTO BOOK-IN AREA WITH POLICE RADIO "ON".

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz

GORDON L. FRANZ

Signature and Printed Name

02/27/2017

Date

80-003945

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JAN 25 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 11:27

Date of Inspection: 01/24/2017

Serial Number: 80-003945
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22315080A2 Exp: 09/05/2017
0.000	0.050	0.079	0.199	0.077
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.081	0.201	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GORDON L FRANZ

Signature and Printed Name

01/24/2017
Date