

5076 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 07:54

Date of Inspection: 11/22/2017

Serial Number: 80-005076
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09216080A2 Exp: 05/05/2018
0.000	0.047	0.076	0.196	0.000 / 0.000
0.000	0.047	0.077	0.196	0.000 / 0.000
0.000	0.048	0.077	0.197	0.000 / 0.000

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance GAS TUBE NOT CONNECTED

Non-compliance: GAS TUBE NOT CONNECTED.

*- GAS TUBE NOT CONNECTED DURING DRY GAS TEST.
- RECONNECTED & RETESTED*

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DAVID M. FELTOVIC 1965

MIKE FELTOVIC

Signature and Printed Name

11/22/2017
Date