



RECEIVED

NOV 24 2017

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO  
Time of Inspection: 07:43

Date of Inspection: 11/24/2017

Serial Number: 80-005249  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201603D<br>Exp: 03/08/2018 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201604C<br>Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:17316080A2<br>Exp: 08/05/2018 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.047   | 0.077   | 0.194   | 0.081   |
| 0.000                      | 0.047   | 0.077   | 0.196   | 0.081   |
| 0.000                      | 0.047   | 0.078   | 0.195   | 0.081   |

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded RESIDUAL SOLUTION FROM PREV.TEST .

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Sgt B Collito 1193*

BRAD R COLLITO

Signature and Printed Name

11/24/2017  
Date

80-005249 *QAS*

RECEIVED

OCT 08 2017

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

 Agency: HERNANDO COUNTY SO  
 Time of Inspection: 20:02

Date of Inspection: 10/08/2017

 Serial Number: 80-005249  
 Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201603D<br>Exp: 03/08/2018 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201604C<br>Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:17316080A2<br>Exp: 08/05/2018 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.047   | 0.077   | 0.195   | 0.079   |
| 0.000                      | 0.047   | 0.077   | 0.195   | 0.080   |
| 0.000                      | 0.047   | 0.078   | 0.196   | 0.080   |

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BRAD R. COLLITO

Signature and Printed Name

10/08/2017  
 Date

80-005249

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED  
JAN 19 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO  
Time of Inspection: 10:11

Date of Inspection: 01/19/2017

Serial Number: 80-005249  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201507A<br>Exp: 07/14/2017 | 0.08g/210L Test (g/210L)<br>Lot#:201601F<br>Exp: 01/26/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201505A<br>Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:12716080A7<br>Exp: 06/05/2018 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.045   | 0.077   | 0.196   | 0.080   |
| 0.000                      | 0.046   | 0.077   | 0.199   | 0.081   |
| 0.000                      | 0.046   | 0.076   | 0.199   | 0.080   |

Number of Simulators Used: 5

Remarks:

A F / M A: Ambient Fail. Non-compliance: AMBIENT FAIL.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BRAD R COLLITO

Signature and Printed Name

01/19/2017  
Date

80-005249

RECEIVED  
JAN 18 2017  
FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO  
Time of Inspection: 08:40

Date of Inspection: 01/19/2017

Serial Number: 80-005249  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect |     | No |
| Diagnostic Check (Post-Inspection): OK      |     | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: Exp: | 0.08g/210L Test (g/210L)<br>Lot#: Exp: | 0.20g/210L Test (g/210L)<br>Lot#: Exp: | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: Exp: |
|----------------------------|--|--|--|---|
| / 0.000                    |  |  |  |   |
|                            |  |  |  |   |
|                            |  |  |  |   |

Number of Simulators Used: 5

Remarks:  
00: Ambient Fail, Ambient Fail. Non-compliance: ODOR IN ROOM CAUSING AMBIENT AND PURGE FAIL .

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Sgt B Collito*

BRAD R COLLITO  
Signature and Printed Name

01/19/2017  
Date