

6/16/17 SP

RECEIVED

# Florida Department of Law Enforcement JUN 01 2017

## Alcohol Testing Program

FDLE  
Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO  
Time of Inspection: 13:54

Date of Inspection: 05/26/2017

Serial Number: 80-006169  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06616080A2 Exp: 04/05/2018
0.000	0.050	0.074 / 0.078	0.200	0.078
0.000	0.051	0.075 / 0.079	0.201	0.078
0.000	0.051	0.076 / 0.079	0.201	0.077

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.SIM CHANGED,WARM UP REPEAT

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D/S C. [Signature] 5/11/1607

CHRISTOPHER L THOMAS

Signature and Printed Name

05/26/2017  
Date