

RECEIVED

Florida Department of Law Enforcement  
Alcohol Testing Program

MAR 06 2017

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO  
Time of Inspection: 11:04

Date of Inspection: 03/04/2017

Serial Number: 80-006471  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 562838 Exp: 09/16/2019
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.077	0.199	0.078
0.000	0.048	0.077	0.200	0.078

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded.

DURING 1<sup>ST</sup> AF/MA, I HAD RESIDUAL MOUTH ALCOHOL SOLUTION IN MY MOUTH, CAUSING RANGE EXCEEDED. I RINSED MY MOUTH WITH WATER AND WAITED 20 MINUTES. RE-RAN AF/MA WITH "SLOPE NOT MET". JB

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

8568/114

JON B HUFF

Signature and Printed Name

03/04/2017  
Date