

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

AUG 30 2017

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FHP

Time of Inspection: 08:30

Date of Inspection: 08/30/2017

Serial Number: 80-006628

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:

BYPASS AI TO OPERATE INSTRUMENT

QJm

\* NOT a compliance check \*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

08/30/2017  
Date

8/30/17  
JL