



# Fort Walton Beach Police Department

---

---

## Memorandum

TO: Patrick Murphy, Ph.D., Department Inspector

FROM: Sgt. James Martin, Agency Inspector

COPIES TO: n/a

SUBJECT: Intoxilyzer 8000 Serial #80-000738

DATE: December 7, 2017

---

The Fort Walton Beach Police Department has purchased and placed a brand new Intoxilyzer 8000 into service; our previous Intoxilyzer 8000 serial #80-000738 is being returned to us from CMI, Inc. This instrument requires extensive repairs, this agency is not inclined to spend money on, before it can be returned to service.

Please remove the instrument from the list of active Intoxilyzers.

**Murphy, Patrick**

---

**From:** Robert Hull <rhull@fwb.org>  
**Sent:** Wednesday, February 01, 2017 7:25 AM  
**To:** Murphy, Patrick  
**Subject:** Loaner intox

Pat,

Last night while doing an Agency Inspection, I injured myself during the inspection. It went past midnight into February due to me getting medical attention. Upon trying to finish the inspection, the connectors sent for the new simulators were set up incorrectly and pumped solution into the sample chamber. I am going to box up the Intox and send it to CMI for an estimate to do the repair.

Is it possible we can get a loaner Intox while ours is out of service?

Thank you,

Sgt. Steve Hull  
C "Charlie" Squad Supervisor - Uniform Patrol Division  
Fort Walton Beach Police Department  
7 Hollywood Blvd. NE  
Fort Walton Beach, FL 32548

(850) 833-9900 (Non-Emergency, 24-7)  
(850) 833-9684 (Fax)  
(850) 259-8848 (Mobile)