

951
SP

Platt, Shayla

From: Christopher Norwood <cnorwood@hendrysheriff.org>
Sent: Sunday, November 19, 2017 11:59 PM
To: Platt, Shayla
Subject: November 2017 Inspections
Attachments: November 2017 Intox Inspections.pdf

Shayla,

I have attached the November 2017 inspections. I did have an issue where I used too much mouth alcohol and received a range exceeded reading. I repeated the test with less mouth alcohol and had no further issues. Thanks.

Christopher J. Norwood
Sergeant #137
Hendry County Sheriff's Office
101 S. Bridge St.
Labelle, FL 33935
863-674-5600

In a Republic that honors the core ideas of democracy, the greatest amount of power is given to those called the Guardians. Only those with the most impeccable character are chosen to bear the responsibility for protecting the democracy. - Plato

Private and Confidential: The sender intends to communicate the contents of this transmission only to the person to whom it is addressed. This transmission may contain information that is privileged, confidential or otherwise exempt from disclosure under applicable law. If you have received this transmission in error, please notify us immediately and promptly destroy the original transmission. Please notify the sender of this error by replying to e-mail. Under Florida law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing. Telephone contact number is (863) 674-5600.