

**Bell, Danielle**

---

**From:** Dan Lyons <dlyo@OSCEOLA.ORG>  
**Sent:** Saturday, December 02, 2017 3:31 PM  
**To:** Bell, Danielle; 'McKnight, Jamie (JMcknight@sao9.org)'; 'Atkinson, Karsona'; 'Labbe, Linda'  
**Cc:** George Maroney; John Haydel  
**Subject:** Monthly Intox inspections Dec 2017 Osceola CO SO  
**Attachments:** DEC 1 2017.pdf

ALCON,

Agency Inspection Report, conducted 12/01/2017 on the following instruments;  
80-000966, 80-000968, 80-001715, 80-003935, 80-003937 , 80-004587 and 80-005935, .

Ref to instrument 80-000966, Interferent Detect Test: Int det, Simulator adjusted, then re-tested. The re-test results are within standards. The inspection was not Re-Started.

Ref to instrument 80-003937, Int Det; RFI Detected. Deputy with a cell phone received a call, RFI Detected. Subject advised cell px off, to leave the room. , then re-tested. The re-test results are within standards. The inspection was not Re-started. 0.08 g/210L Dry Gas Std Test. Ambient Fail – Failed to reconnect the tube. Tube reconnected , then re-tested. The re-test results are within standards. The inspection was not Re-started

All intox inspections are complete and in compliance with 11D-8.

I will be in Hawaii from Dec 4<sup>th</sup> –Dec 22<sup>nd</sup>, still available for all questions and concerns. I will check my email periodically or contact me via my Cell Phone. Just be considerate of the -5 hour time difference if my response is not immediate.

Thank you, have a Safe Holiday.

V/R

Dan Lyons  
Osceola County Sheriff's Office  
Spec Ops/Traffic/DUI Enforcement Unit  
Agency 407-348-1100  
Agency Cell 407-269-1980  
Cell 321-960-2066

---

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

**Bell, Danielle**

---

**From:** Dan Lyons <dlyo@OSCEOLA.ORG>  
**Sent:** Wednesday, July 12, 2017 9:31 PM  
**To:** Bell, Danielle  
**Subject:** RE: Monthly Agency Inspection Report 07/11/17

Good Morning,

For the first two, 968 and 3937 inspections, Done

The aborted the test during the slope not met section, was an error on my part. I should have hit repeat and I did not.

I forgot to use the mouth alcohol solution and should have hit repeat.

Why did 3 instruments not detect slope on the first shot?

The third 3935, I restarted as instructed after blowing SNM for MA and I got an actual reading instead of SNM. So I grabbed someone nearby and had them blow the alcohol free breath for me.

This one I blew the and did not get SNM, instead I got a reading of .011 and I started again and still not a reading of .04, so I grabbed some nearby and had them blow for me for the alcohol free breath and added more mouth alcohol solution.

I have attached the changes..

Thanks,

V/R,

Dan Lyons  
Osceola County Sheriff's Office  
Spec Ops/Traffic/DUI Enforcement Unit  
Agency 407-348-1100  
Agency Cell 407-269-1980  
Cell 321-960-2066

---

**From:** Bell, Danielle [<mailto:DanielleBell@fdle.state.fl.us>]  
**Sent:** Wednesday, July 12, 2017 4:51 PM  
**To:** Dan Lyons <dlyo@OSCEOLA.ORG>  
**Subject:** RE: Monthly Agency Inspection Report 07/11/17

Good afternoon,

I just got a second to look at the inspections for this month.

On the first 968 and 3937 inspections, please mark them as "does not comply" and single line through the complies since those inspections at 14:04 & 14:05 were not completed. Please indicate why you aborted the test during the slope

not met section. Did you have to change mouth alcohol bottles? Why did 3 instruments not detect slope on the first shot? Initial and date the changes and send me a copy.

You have a note on 3935 that the intox did not meet snm, but the inspection indicates that it did. Was the alcohol free / mouth alcohol test repeated on this inspection and you backed up through the AF / MA note that the Intox generated? If that is the case, please indicate that the AF/MA was repeated due to "your already noted comments." and indicate what you did to get the Intox to detect the mouth alcohol? (ie: changed mouth alcohol bottles, used more mouth alcohol, etc.)

Thanks,

*Danielle*

*Danielle M. Bell*

Department Inspector  
Alcohol Testing Program  
Florida Department of Law Enforcement  
(850) 617-1286 (office)  
(850) 274-2757 (cell)  
(850) 921-3787 (fax)

---

**From:** Dan Lyons [<mailto:dlyo@OSCEOLA.ORG>]  
**Sent:** Tuesday, July 11, 2017 5:08 PM  
**To:** 'McKnight, Jamie ([JMcknight@sao9.org](mailto:JMcknight@sao9.org))'; 'Atkinson, Karsona'; 'LindaLabbe@flhsmv.gov'; Bell, Danielle  
**Cc:** Daniel Marquith  
**Subject:** Monthly Agency Inspection Report 07/11/17

ALCON,  
Agency Inspection Report, conducted 05/24/2017 on the following instruments;

80-000968,  
80-003935,  
80-003937,  
80-004587 and  
80-005935.

V/R,

Dan Lyons  
Osceola County Sheriff's Office  
Spec Ops/Traffic/DUI Enforcement Unit  
Agency 407-348-1100  
Agency Cell 407-269-1980  
Cell 321-960-2066

---

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution or copy of this communication is strictly prohibited. If you have received this communication in error, please contact the sender by reply E-mail and destroy all copies of the original message.

---

Attention: The information contained in this E-mail message is privileged and confidential information intended only for the use of the individual(s) named above. If