

Bell, Danielle

80-001042
80-001732

From: Lawrence Kolbicka <lkolbicka@pinellas-park.com>
Sent: Friday, July 14, 2017 12:14 PM
To: Bell, Danielle
Subject: Inspection issues...

Hi Danielle,

I had two issues with my inspections that were my fault.

80-001042

One was that i didnt use enough MA on the MA test and it read out of range. i redid it and it worked fine.

Second was on the dry gas. I forgot to hook up the hose until the first test was almost over... I redid it and there were no issues.

80-001732

Everything is fine and I just wanted you to be aware of them when you saw them

Thank You,

Larry

--
OFC Lawrence R Kolbicka II
Special Operations Unit
Traffic Homicide Investigator
Traffic Enforcement

Pinellas Park Police Department

Bell, Danielle

From: Lawrence Kolbicka <lkolbicka@pinellas-park.com>
Sent: Wednesday, May 03, 2017 12:54 PM
To: Bell, Danielle
Subject: RFI

Follow Up Flag: Follow up
Flag Status: Flagged

Danielle,

While conducting a control test on the water wet bath, i got an RFI detect because i had my radio on and when I put my hand on my hip, i hit the transmit button. I placed my radio in another room, ran the test again and had no issues.

Larry

--

OFC Lawrence R Kolbicka II
Special Operations Unit
Traffic Homicide Investigator
Traffic Enforcement

Pinellas Park Police Department