

80-001127
80-001055

Bell, Danielle

From: Wilson, Barbie <BWilson@polksheriff.org>
Sent: Thursday, February 09, 2017 11:26 AM
To: Bell, Danielle
Subject: FW: Scan Job from Copier
Attachments: scan.pdf

Good Morning,

I just wanted to advise you that Instrument 80-001127 failed the .20 inspection the first time and the error message was Control Outside Tolerance; however, after performing the test the second time it passed. I took it out of service it any way and replaced it with another instrument for use. Also, did you receive Instrument 80-001055 from Enforcement Electronics because I never received it back from having it serviced. I will give them a call today.

Thank you,

Wilson