

Bell, Danielle

From: Steve Gassen <sgassen@tspd.us>
Sent: Friday, April 21, 2017 10:11 AM
To: Bell, Danielle
Subject: April Inspection
Attachments: April 2017 Inspection.pdf

Danielle,
I conducted this month's inspection and there was a leak in the house on my 0.08 sample. I replaced the hose and had a successful test.

Also I need to discuss an unrelated item with you when you can call me at 727-919-7648

Thank you

STEVE GASSEN
POLICE OFFICER.



TARPON SPRINGS
POLICE DEPARTMENT

"Building a Better Future through Excellence in Policing"

444 South Huey Avenue
Tarpon Springs, FL 34689
(727) 938-2849 – Main
(727) 939-4546 - Direct
(727) 937-0193 – Fax
sgassen@tspd.us



Please Note: The State of Florida has a very broad public records law under chapter 119. Most written material to or from City officials regarding City business are public records and available to the public or media upon request. Your e-mail communications and/or e-mail attachments may be subject to public disclosure.

Bell, Danielle

From: Steve Gassen <sgassen@tspd.us>
Sent: Tuesday, January 31, 2017 4:41 PM
To: Bell, Danielle
Subject: Re: Couple of things

I just did the second one and it is in compliance. I am sending my sim off for repair

Thank you

Officer Steve Gassen
Tarpon Springs Police
444 S Huey Ave
Tarpon Springs, FL
34689

Sent via mobile device

On Tue, Jan 31, 2017 at 4:22 PM -0500, "Bell, Danielle" <DanielleBell@fdle.state.fl.us> wrote:

Yep! That is perfect. As long as you know what the issue is and I am notified, you have met the rule requirement. As long as you still have 2 simulators, I see no reason not to continue with another inspection.

Thanks,

Danielle

Danielle M. Bell
Department Inspector
Alcohol Testing Program
Florida Department of Law Enforcement
(850) 617-1286 (office)
(850) 274-2757 (cell)
(850) 921-3787 (fax)

From: Steve Gassen [<mailto:sgassen@tspd.us>]
Sent: Tuesday, January 31, 2017 4:02 PM
To: Bell, Danielle
Subject: Couple of things

Danielle,
Can you send me Alcohol and Acetone samples please?

Also my January 31, 2017 inspection failed because my simulator heater failed and I had to take it out of service and will be doing the inspection again minus one simulator?

Is this the kind of notification you wanted?

STEVE GASSEN

POLICE OFFICER



**TARPON SPRINGS
POLICE DEPARTMENT**

"Building a Better Future through Excellence in Policing"

444 South Huey Avenue

Tarpon Springs, FL 34689

(727) 938-2849 – Main

(727) 939-4546 - Direct

(727) 937-0193 – Fax

sgassen@tspd.us



Please Note: The State of Florida has a very broad public records law under chapter 119. Most written material to or from City officials regarding City business are public records and available to the public or media upon request. Your e-mail communications and/or e-mail attachments may be subject to public disclosure.