

SP

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Thursday, August 17, 2017 6:17 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - August 17
Attachments: Aug 17.pdf

Please see attached documents for this month's inspection reports.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org



1266
SP

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Wednesday, July 12, 2017 3:25 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - July 2017
Attachments: July 17.pdf

Please see attached documents for this month's inspection reports.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org



1266
SP

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Saturday, June 24, 2017 6:19 PM
To: TWELLS@sa18.org; Darin Morgan; Melbourne BAR; Platt, Shayla; Lisa Weggeland
Subject: Palm Bay PD Intoxilyzer Documents - June
Attachments: SKMBT_C554e17062413100.pdf; ATT00001.txt

Follow Up Flag: FollowUp
Flag Status: Flagged

Please see attached;

SP 1266

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Sunday, June 04, 2017 10:21 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - May 17
Attachments: May 17.pdf

Sorry for the delay; please see attached documents for this month's inspection reports.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org



1266
SP

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Friday, May 05, 2017 3:35 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - 80-001266 Dept Inspection
Attachments: 80-001266 - 2017.pdf

Please see attached documents for the department inspection for 2017 for 1266.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org



INSTRUMENT PROCESSING SHEET

Agency Palm Bay PD S/N 80-001266
Date In 4/27/17 Date Out 5/1/17 Ship P/U H/D CMI EE

RECEIVED
MAY 04 2017
FDLE
Alcohol Testing Program

Intake Performed By SP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: One foot in back is loose, also some foot is attached with wrong screw

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 145
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm .144 (.139 - .169)
 36mm .164 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .492 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G3709	201603D 3-8-18
0.08	DR1279	201611B 11-15-18
0.20	DR3856	201604C 4-5-18
0.08 DGS	N/A	AG626605 9-22-18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure
 ID# 28427 1014 Gauge
1012 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G3709
0.08	DR1279
0.20	DR3856

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other DIAGNOSTIC CHECKS

Notes: RAN ADDITIONAL DIAGNOSTIC CHECKS TO ENSURE ACCURACY & RELIABILITY. SP

QC

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

JJ Graham
Quality Control Review

5/4/17 Date

I HEREBY CERTIFY THE AUTHENTICITY OF THE SEAL AFFIXED HERETO.

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 04 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Time of Inspection: 11:46

Date of Inspection: 05/01/2017

Serial Number: 80-001266

Software: 8100.27

BAC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.048	0.078	0.198	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079

Standard Deviations	0.0004	0.0005	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

DAS
JE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

05/01/2017
Date

I HEREBY CERTIFY THE AUTHENTICITY OF THE SEAL AFFIXED HERETO.

SP

STABILITY CHECKS - INSTRUMENT # 80-001266 - PALM BAY P.D. SP

PALM BAY P.D.
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-001266
05/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.007	09:37
Air Blank	0.000	09:38
Control Test	0.048	09:38
Air Blank	0.001	09:39
Control Test	0.148	09:40
Air Blank	0.000	09:40
Control Test Stats	0.0477	09:40
Average	0.0006	
Std Dev	0.2112	
Rel. Std Dev(%)	0.7466	

SP

Operator's Signature

PALM BAY P.D.
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-001266
05/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.077	09:42
Air Blank	0.000	09:42
Control Test	0.078	09:43
Air Blank	0.000	09:44
Control Test	0.077	09:44
Air Blank	0.000	09:45
Control Test Stats	0.0773	09:45
Average	0.0006	
Std Dev	0.7466	
Rel. Std Dev(%)	0.7466	

SP

Operator's Signature

PALM BAY P.D.
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-001266
05/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.195	09:32
Air Blank	0.000	09:33
Control Test	0.197	09:34
Air Blank	0.000	09:34
Control Test	0.197	09:35
Air Blank	0.000	09:35
Control Test Stats	0.1957	09:35
Average	0.0006	
Std Dev	0.2086	
Rel. Std Dev(%)	0.2936	

SP

Operator's Signature

PALM BAY P.D.
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-001266
05/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.060	09:46
Air Blank	0.000	09:47
Control Test	0.080	09:47
Air Blank	0.000	09:48
Control Test	0.079	09:48
Air Blank	0.000	09:49
Control Test Stats	0.0797	09:49
Average	0.0006	
Std Dev	0.2006	
Rel. Std Dev(%)	0.7247	

DAS

SP

Operator's Signature

I HEREBY CERTIFY THE AUTHENTICITY OF THE SEAL AFFIXED HERETO.

SP

DIAGNOSTIC CHECKS - INSTRUMENT#
PALM BAY P.D. 80-001266

PALM BAY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 80-001266
05/03/2017
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

PALM BA
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 80-001266
05/03/2017
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

PALM BA
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 80-001266
05/03/2017
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

RWB

JE

I HEREBY CERTIFY THE AUTHENTICITY
OF THE SEAL AFFIXED HERETO.

SP

SP 1266

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Sunday, February 26, 2017 1:25 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - Feb '17
Attachments: Feb 17.pdf

Please see attached documents for this month's inspection reports.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org

