


1266 SP
Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Tuesday, December 26, 2017 6:28 PM
To: Darin Morgan; Rachel Kunerth; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - December 17
Attachments: Dec 17.pdf

Please see attached documents for this month's inspection reports.

Edwin J. Lutz III

Career Police Officer
Drug Recognition Expert
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907



CONFIDENTIALITY NOTICE:

The information contained in this e-mail is strictly confidential and prepared solely for the use of the intended recipient(s). The copyright of this communication belongs to the City of Palm Bay. If you are not the intended recipient, please do not read, use, disseminate, distribute or copy this message or attachments. If you have received this message in error, please notify the sender immediately and delete this message. Views expressed in this message are those of the individual sender, and are not necessarily the views of the City of Palm Bay City Council.

DISCLAIMER:

Before opening any attachments, please check them for viruses and defects. The sender does not accept liability for any viruses, errors or omissions in the contents of this message or attachment, which arise as a result of email transmission.

PUBLIC RECORDS NOTICE:

All e-mail sent to and received from the City of Palm Bay, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.

Promoting a Green Palm Bay - Please consider the environment before printing this e-mail!


1266 SP
Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Friday, December 01, 2017 2:01 PM
To: Darin Morgan; Rachel Kunerth; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records November 17
Attachments: Nov 17.pdf

Please see attached documents for this month's inspection reports.

Edwin J. Lutz III

Career Police Officer
Drug Recognition Expert
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907



CONFIDENTIALITY NOTICE:

The information contained in this e-mail is strictly confidential and prepared solely for the use of the intended recipient(s). The copyright of this communication belongs to the City of Palm Bay. If you are not the intended recipient, please do not read, use, disseminate, distribute or copy this message or attachments. If you have received this message in error, please notify the sender immediately and delete this message. Views expressed in this message are those of the individual sender, and are not necessarily the views of the City of Palm Bay City Council.

DISCLAIMER:

Before opening any attachments, please check them for viruses and defects. The sender does not accept liability for any viruses, errors or omissions in the contents of this message or attachment, which arise as a result of email transmission.

PUBLIC RECORDS NOTICE:

All e-mail sent to and received from the City of Palm Bay, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.

Promoting a Green Palm Bay - Please consider the environment before printing this e-mail!

1266
SP


Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Friday, October 27, 2017 10:35 PM
To: Darin Morgan; Rachel Kunerth; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - Oct 2017
Attachments: Oct 17.pdf

Please see attached documents for this month's inspection reports.

Edwin J. Lutz III

Career Police Officer
Drug Recognition Expert
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907



CONFIDENTIALITY NOTICE:

The information contained in this e-mail is strictly confidential and prepared solely for the use of the intended recipient(s). The copyright of this communication belongs to the City of Palm Bay. If you are not the intended recipient, please do not read, use, disseminate, distribute or copy this message or attachments. If you have received this message in error, please notify the sender immediately and delete this message. Views expressed in this message are those of the individual sender, and are not necessarily the views of the City of Palm Bay City Council.

DISCLAIMER:

Before opening any attachments, please check them for viruses and defects. The sender does not accept liability for any viruses, errors or omissions in the contents of this message or attachment, which arise as a result of email transmission.

PUBLIC RECORDS NOTICE:

All e-mail sent to and received from the City of Palm Bay, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.

Promoting a Green Palm Bay - Please consider the environment before printing this e-mail!

SP

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Thursday, August 17, 2017 6:17 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - August 17
Attachments: Aug 17.pdf

Please see attached documents for this month's inspection reports.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org



1266
SP

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Wednesday, July 12, 2017 3:25 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - July 2017
Attachments: July 17.pdf

Please see attached documents for this month's inspection reports.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org



1266
SP

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Saturday, June 24, 2017 6:19 PM
To: TWELLS@sa18.org; Darin Morgan; Melbourne BAR; Platt, Shayla; Lisa Weggeland
Subject: Palm Bay PD Intoxilyzer Documents - June
Attachments: SKMBT_C554e17062413100.pdf; ATT00001.txt

Follow Up Flag: FollowUp
Flag Status: Flagged

Please see attached;

Platt, Shayla

SP 1266

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Sunday, June 04, 2017 10:21 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - May 17
Attachments: May 17.pdf

Sorry for the delay; please see attached documents for this month's inspection reports.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org



1266
SP

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Friday, May 05, 2017 3:35 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - 80-001266 Dept Inspection
Attachments: 80-001266 - 2017.pdf

Please see attached documents for the department inspection for 2017 for 1266.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org



INSTRUMENT PROCESSING SHEET

Agency Palm Bay PD

S/N 80-001266

Date In 4/27/17

Date Out 5/1/17

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

- Intake** Performed By SP
- ☐ Registration
 - ☒ Annual
 - ☐ Return from CMI
 - ☐ Return from Enforcement
 - ☐ Electronics
 - ☐ Other

Visual Inspection:

- OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:

- ☐ Power cord
- ☐ Printer Cable
- ☒ Other Static Bag

Notes: One foot in back is loose. Also, same foot is attached with wrong screw

- Quality Checks** Performed By SP
- ☒ Breath Tube Screen
 - ☒ Replace O-Rings
 - ☒ Instrument Set Up Verified
 - ☒ R-Value 145
 - ☒ Flow Verification (L/s)

Flow Column # ATP103
 32mm .144 (.139 - .169)
 36mm .164 (.156 - .190)
 53mm .234 (.228 - .278)
 103mm .492 (.447 - .547)

- ☒ Barometric Pressure Check
- Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G3709	201603D 3-8-18
0.08	DR1279	201611B 11-15-18
0.20	DR3856	201604C 4-5-18
0.08 DGS	N/A	AG626605 9-22-18

- Flow Calibration** Performed By SP
- ☒ Flow Calibration N/A
 - ☐ Flow Calibration Complete
 - Flow Column # Alcohol Testing Program
 - ☐ 5L/min - 17mm
 - ☐ 15L/min - 53mm
 - ☐ 30L/min - 103mm
 - ☐ R-Value
 - ☐ Post Calibration Verification (L/s)

Flow Column #
 32mm (.139 - .169)
 36mm (.156 - .190)
 53mm (.228 - .278)
 103mm (.447 - .547)

Maintenance Performed By SP

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other

Suggested Service

Optical Bench Calibration Performed By SP

- ☒ Optical Bench Calibration N/A
- ☐ Optical Bench Calibration Complete

Barometric Pressure Gauge ID #

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

☐ Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

- ☒ Barometric Pressure ID# 28427 1014 Gauge
- 1012 Instrument

Mouth Alcohol Solution Lot # 2016-A

Acetone Stock Solution Lot # 2017-A

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G3709
0.08	DR1279
0.20	DR3856

Attachments

- ☒ Form 41
- ☒ Pre-Stability Tests
- ☐ Flow Calibration
- ☐ Optical Bench Cal
- ☐ Post-Stability Tests
- ☒ Other DIAGNOSTIC CHECKS

Notes: RAN ADDITIONAL DIAGNOSTIC CHECKS TO ENSURE ACCURACY & RELIABILITY. SP

QC: [Signature]

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Quality Control Review

Date

Revised November 2012

I HEREBY CERTIFY THE AUTHENTICITY OF THE SEAL AFFIXED HERETO.

RECEIVED
 MAY 04 2017
 FDLE
 Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 04 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Time of Inspection: 11:46

Date of Inspection: 05/01/2017

Serial Number: 80-001266

Software: 8100.27

BEC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.048	0.078	0.198	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
Standard Deviations	0.0004	0.0005	0.0005	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

DAS
JB

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D. PLATT

Signature and Printed Name

05/01/2017
Date

I HEREBY CERTIFY THE AUTHENTICITY
OF THE SEAL AFFIXED HERETO.

SP

STABILITY CHECKS- INSTRUMENT # 80-001266 - PALM BAY P.D.

SP

PALM BAY P.D.
Intoxilizer - Alcohol Analyzer
Model 8000
05/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.000	09:37
Air Blank	0.000	09:38
Control Test	0.000	09:39
Air Blank	0.000	09:39
Control Test	0.000	09:40
Air Blank	0.000	09:40
Control Test	0.000	09:41
Average	0.000	
Std Dev	0.000	
Rel Std Dev(%)	0.000	

PALM BAY P.D.
Intoxilizer - Alcohol Analyzer
Model 8000
05/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.000	09:42
Air Blank	0.000	09:42
Control Test	0.000	09:43
Air Blank	0.000	09:43
Control Test	0.000	09:44
Air Blank	0.000	09:44
Control Test	0.000	09:45
Average	0.000	
Std Dev	0.000	
Rel Std Dev(%)	0.000	

PALM BAY P.D.
Intoxilizer - Alcohol Analyzer
Model 8000
05/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.000	09:32
Air Blank	0.000	09:33
Control Test	0.000	09:34
Air Blank	0.000	09:34
Control Test	0.000	09:35
Air Blank	0.000	09:35
Control Test	0.000	09:36
Average	0.000	
Std Dev	0.000	
Rel Std Dev(%)	0.000	

PALM BAY P.D.
Intoxilizer - Alcohol Analyzer
Model 8000
05/01/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.000	09:46
Air Blank	0.000	09:47
Control Test	0.000	09:47
Air Blank	0.000	09:48
Control Test	0.000	09:48
Air Blank	0.000	09:49
Control Test	0.000	09:49
Average	0.000	
Std Dev	0.000	
Rel Std Dev(%)	0.000	

SP

Operator's Signature

SP

Operator's Signature

SP

Operator's Signature

SP

Operator's Signature

I HEREBY CERTIFY THE AUTHENTICITY
OF THE SEAL AFFIXED HERETO.

SP

DIAGNOSTIC CHECKS - INSTRUMENT#
PALM BAY P.D. 80-001266

PALM BAY P.D.
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 80-001266
05/03/2017
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

PALM BA
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 80-001266
05/03/2017
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

PALM BA
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 80-001266
05/03/2017
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

RWB

JE

I HEREBY CERTIFY THE AUTHENTICITY
OF THE SEAL AFFIXED HERETO.

SP

Platt, Shayla

From: Edwin J. Lutz <edwin.lutz@palmbayflorida.org>
Sent: Wednesday, April 19, 2017 10:17 PM
To: Platt, Shayla
Cc: Darin Morgan
Subject: Diagnostic Fail - Analytical Stability Test

Shayla,

Ever see an instrument display this or any idea why it would occur? This instrument is up for its department inspection anyway and I haven't been able to replicate it again, every diagnostic check has passed since. Let me know if you have any ideas why this occurred. Thanks!

[illegible]

SP, 1266

Platt, Shayla

From: Edwin Lutz <Edwin.Lutz@palmbayflorida.org>
Sent: Sunday, February 26, 2017 1:25 PM
To: Darin Morgan; Karen Delmarco; Melbourne BAR; Platt, Shayla; TWELLS@sa18.org
Subject: Palm Bay P.D. Intoxilyzer Records - Feb '17
Attachments: Feb 17.pdf

Please see attached documents for this month's inspection reports.

C.P.O. Edwin J. Lutz # 197

Drug Recognition Expert
A.T.P. Agency Inspector
D.U.I. Enforcement
Palm Bay Police Department
130 Malabar Road, South East
Palm Bay, Florida, 32907
edwin.lutz@palmbayflorida.org

