

1341
SP

Platt, Shayla

From: Joshua Pitts <joshuapitts@desotosheriff.org>
Sent: Friday, September 22, 2017 10:55 PM
To: Platt, Shayla
Subject: DCSO Intoxilyzer Agency Inspections
Attachments: Agency Inspection 09-22-17 1341.pdf; Agency Inspection 09-22-17 1340.pdf

Ma'am,

Attached is the Agency Inspections for both of our instruments. Our 80-001340 instrument has progressively been showing lower readings than our 1341 instrument. Using the same solution from same batch. The 1340 instrument could not pass the inspection this month and it will not be utilized. I will be getting with my captain to have it shipped out for maintenance.

Thank you,

Lieutenant Pitts, Joshua 34J06
DeSoto County Sheriff's Office Jail
208 East Cypress St.
Arcadia FL, 34266
(863)-993-4700 EXT: 2260
(863)-491-6760
joshuapitts@desotosheriff.org

Please Note:

Under Florida law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.