From:

Simken, Robert <SimkenR@ci.eustis.fl.us>

Sent:

Monday, May 08, 2017 10:05 AM

To:

Bell, Danielle

Subject:

Eustis PD

The acetone was fine. It looks like the lid was not tight enough.

When I went back in, the instrument finally showed "Repeat" Must not have waited long enough.

Robert

Detective Robert Simken Senior Officer, E34 Criminal Investigations Division 51 E. Norton Avenue Eustis, FL 32726 Office: (352) 483-5400 X5202

Cell: (352) 267-2157 simkenr@ci.eustis.fl.us



From:

Simken, Robert <SimkenR@ci.eustis.fl.us>

Sent:

Monday, April 03, 2017 8:12 AM

To:

Bell, Danielle

Subject:

Eustis PD March Inspection correction

**Attachments:** 

March 2017 Eustis Police Department Intoxilyzer 8000 Agency Inspection Not in

Compliance due to mouth alcohol.pdf; March 2017 Eustis Police Department Intoxilyzer

8000 Agency Inspection.pdf

Follow Up Flag:

Follow up

Flag Status:

Flagged

Danielle,

Attached is the corrected form 40 where Grieves had to restart the inspection and needed to initial the form showing it was not in compliance.

Robert

Detective Robert Simken Senior Officer, E34 Criminal Investigations Division 51 E. Norton Avenue Eustis, FL 32726 Office: (352) 483-5400 X5202

Cell: (352) 267-2157

simkenr@ci.eustis.fl.us





From:

Simken, Robert <SimkenR@ci.eustis.fl.us>

Sent:

Tuesday, March 28, 2017 7:55 AM

To:

Bell, Danielle

Subject:

Eustis PD Intoxilyzer

**Attachments:** 

DOC032817.pdf

Importance:

High

Good morning Danielle,

I was out for a week with the flu. Matt Grieves did the March inspection for me yesterday. He advised he had introduced the mouth alcohol on the alcohol free sample. Realizing what he did he restarted the inspection and completed a new one.

He did print out both inspections. I am attaching them for your review. What needs to be done or was this acceptable?

#### Robert

Detective Robert Simken Senior Officer, E34 Criminal Investigations Division 51 E. Norton Avenue Eustis, FL 32726 Office: (352) 483-5400 X5202

Cell: (352) 267-2157

cell: (352) 267-2157 simkenr@ci.eustis.fl.us




I cliscussed this w/ Det. Simken.

Starting over from scratch is 100% acceptable.

I indicated that initials were needed is

that the 1st isospection was not compliant.



From:

Simken, Robert <SimkenR@ci.eustis.fl.us>

Sent:

Tuesday, January 24, 2017 9:38 AM

To:

(WINTERSPRINGSBAR@FLHSMV.GOV); Andrea Williams; Bell, Danielle; Bell, Danielle;

Grieves, Matthew; Hunsicker, Megan; Simken, Robert

Subject: Attachments: Eustis Police Department Annual Inspection and Back in Service from FDLE

2017 FDLE Annual Inspection.pdf; Back In Service Returned from FDLE for Annual

Inspection January 2017 Eustis Police Department Intoxilyzer 8000 Agency

Inspection.pdf

Importance:

High

Follow Up Flag:

Follow up

Flag Status:

Flagged

Attached are the 2017 FDLE annual inspection and the EPD Back in Service Inspection returning from FDLE.

Have awesome day!!

Detective Robert Simken Senior Officer, E34 Criminal Investigations Division 51 E. Norton Avenue Eustis, FL 32726 Office: (352) 483-5400 X5202

Cell: (352) 267-2157 simkenr@ci.eustis.fl.us

