



*Sheriff Jerry L. Demings*  
**ORANGE COUNTY SHERIFF'S OFFICE**  
TELEPHONE: 407.254.7000 ❖ P.O. BOX 1440, ORLANDO, FLORIDA 32802-1440 ❖ WWW.OCSO.COM

October 19, 2016

**RECEIVED**

**OCT 26 2016**

**FDLE  
Alcohol Testing Program**

Department Inspector Jake Shanahan  
Florida Department of Law Enforcement  
Alcohol Testing Program  
P.O. Box 1489  
Tallahassee, FL 32302-1489

Dear Inspector Shanahan:

This letter is to recap our telephone conversations on September 27, 2016, and October 19, 2016, regarding the incident which occurred after the Agency Inspection conducted on September 14, 2016.

The Agency Inspection of Intoxilyzer 8000, serial number 80-001418, was conducted on September 14, 2016. The instrument complied with all requirements of 11D-8, and passed the inspection.

I placed 80-001418 into one of the testing rooms for evidentiary breath testing use after the Agency Inspection was completed. I was notified via telephone on Saturday, September 24, 2016 that the instrument had gone into Disabled Mode due to its memory being full. I had previously uploaded instrument 80-001418 on Thursday, September 22, 2016. I returned to work on Tuesday, September 27, 2016, and uploaded 80-001418 on that date to see if this would clear the memory. The instrument still indicated that it was in Disabled Mode due to its memory being full. I removed the instrument from the testing room.

I am unable to perform an Agency Inspection at this time due to the fact that the instrument goes from Standby Mode to Disabled Mode and will not allow me to access the Agency Inspection function. Per your instructions, I am removing the instrument from service and shipping 80-001418 to your attention at FDLE/Alcohol Testing Program.

If you have any questions, please contact me at (407) 836-0800.



*The First Law Enforcement Agency  
in Orange County to Receive Both  
International and State Accreditation*



RECEIVED


OCT 26 2016

FDLE  
Alcohol Testing Program

Department Inspector Jake Shanahan  
October 19, 2016  
Page 2

Sincerely,

Jerry L. Demings  
Sheriff of Orange County

A handwritten signature in cursive script, appearing to read "Kelly Melville".

Agency Inspector Kelly Melville  
Special Operations Division  
Traffic/DUI Testing Center

KM/kmm