

5076
SP

Platt, Shayla

From: Michael Feltovic <Michael.Feltovic@sarasotasheriff.org>
Sent: Friday, November 24, 2017 6:58 AM
To: Platt, Shayla
Subject: Agency Inspection
Attachments: 80-005076 Nov AI.pdf

Shayla,

I have attached an agency inspection report for instrument 80-005076. During the inspection, I forgot to connect the gas tube during the Dry Gas Std test. I reattached the gas tube and conducted another inspection with no issues. I noted this on the report.

Thanks!

Mike



Mike Feltovic

Deputy, Special Operations Bureau

2071 Ringling Blvd

Sarasota, FL

Desk Phone: 941.861.4076

Cell Phone: 941.706.6943


michael.feltovic@sarasotasheriff.org



Michael Feltovic

5076

Law Enforcement Deputy
Spec Ops-Traffic
6010 Cattleridge Blvd
Sarasota, Florida 34232
Phone Number: (941) 706-6943
Michael.Feltovic@sarasotasheriff.org
SarasotaSheriff.org



"Please Note: All mail to and from the Sarasota County Sheriff's Office is subject to the Public Records Law of Florida."