

Bell, Danielle

80-005249 QMS

From: Collito, Brad <bcollito@hernandosheriff.org>
Sent: Thursday, January 19, 2017 10:21 AM
To: Bell, Danielle
Subject: Instrument 5249

Follow Up Flag: Follow up
Flag Status: Flagged

Hi Danielle,

I just wanted to follow my voicemail up with an email. I'm pretty sure that I was just rambling on. Instrument 5249 had an Ambient Fail and a Purge fail. I waited an hour after turning on a fan to evacuate any type of odor in the room and conducted the inspection again. The instrument just had an Ambient Fail. I have not uploaded the results yet. I wanted to talk with you first since the instrument was not in compliance for two inspections. I was not 100% sure how to proceed since the instrument was noncompliant for 2 inspections back to back.

Thanks

Brad

Sergeant Brad Collito
Intake/ Release Unit
Hernando County Detention Center
16425 Spring Hill Dr
Brooksville, FL 34604
T(352)544-2334 F(352)544-2349

bcollito@hernandosheriff.org<mailto:bcollito@hernandosheriff.org>

Please note: Florida has a very broad Public Records Law. Most written communications to or from the Hernando County Sheriff's Office are public records available to anyone upon request. Your e-mail, including your e-mail address, may therefore, be subject to public disclosure.