

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUN 01 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 07:50

Date of Inspection: 05/24/2017

Serial Number: 80-000786
Software: 8100.27

BCC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.051	0.081	0.199	0.080
0.000	0.051	0.081	0.199	0.079
0.000	0.051	0.081	0.199	0.079
0.000	0.050	0.081	0.199	0.079
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.079
0.000	0.050	0.081	0.198	0.079
0.000	0.051	0.081	0.199	0.078

Standard Deviations	0.0005	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------


Pom

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

DAVID E REYES-RIVERA

05/24/2017
 Date

JS