

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

MAR 02 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DIXIE COUNTY SO
Time of Inspection: 09:23

Date of Inspection: 03/02/2017

Serial Number: 80-000809
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.048	0.080	0.200	0.000 / 0.080
0.000	0.048	0.080	0.201	0.000 / 0.081
0.000	0.048	0.080	0.201	0.000 / 0.081
0.000	0.049	0.081	0.201	0.000 / 0.080
0.000	0.049	0.080	0.201	0.000 / 0.080
0.000	0.049	0.080	0.201	0.079 / 0.080
0.000	0.049	0.081	0.201	0.080 / 0.081
0.000	0.049	0.080	0.201	0.080 / 0.081
0.000	0.049	0.080	0.201	0.080 / 0.081
0.000	0.049	0.081	0.201	0.080 / 0.081

Standard Deviations	0.0004	0.0004	0.0003	0.0420 / 0.0005
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance DGS NOT CONNECTED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

JAKE L SHANAHAN

03/02/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 10 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DIXIE COUNTY SO
Time of Inspection: 10:58

Date of Inspection: 02/08/2017

Serial Number: 80-000809
Software: 8100.27

JSC
BSK

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/18/2018 03/08/2018 (2)	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG626605 Exp: 09/22/2018
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.200	0.081
0.000	0.049	0.080	0.200	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.201	0.081
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.081

Standard Deviations	0.0005	0.0005	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

JMS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN
Signature and Printed Name

02/08/2017
Date