

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

MAR 14 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO
Time of Inspection: 20:42

Date of Inspection: 03/03/2017

Serial Number: 80-000830
Software: 8100.27

BK

BCC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.052	0.082 / 0.081	0.200	0.079
0.000	0.052	0.089 / 0.082	0.201	0.080
0.000	0.052	0.085 / 0.082	0.202	0.080
0.000	0.052	0.085 / 0.082	0.203	0.080
0.000	0.052	0.085 / 0.083	0.202	0.080
0.000	0.052	0.084 / 0.082	0.202	0.080
0.000	0.052	0.085 / 0.083	0.203	0.079
0.000	0.052	0.084 / 0.083	0.202	0.079
0.000	0.051	0.084 / 0.083	0.203	0.079
0.000	0.051	0.083 / 0.083	0.203	0.079

QPM

Standard Deviations	0.0004	0.0018 / 0.0006	0.0009	0.0005
---------------------	--------	-----------------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance SUSPECT POWER SURGE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

03/03/2017
Date