

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

FEB 10 2017

FDLE

Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CAPE CORAL PD  
Time of Inspection: 12:01

Date of Inspection: 02/02/2017

Serial Number: 80-000859  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.049	0.072 / 0.079	0.194	0.080
0.000	0.049	0.076 / 0.080	0.195	0.080
0.000	0.049	0.075 / 0.080	0.195	0.080
0.000	0.049	0.077 / 0.079	0.195	0.079
0.000	0.049	0.078 / 0.079	0.195	0.080
0.000	0.050	0.077 / 0.079	0.195	0.080
0.000	0.050	0.078 / 0.079	0.195	0.079
0.000	0.050	0.074 / 0.079	0.195	0.080
0.000	0.050	0.078 / 0.080	0.194	0.079
0.000	0.050	0.077 / 0.079	0.194	0.080

Standard Deviations	0.0005	0.0019 / 0.0004	0.0004	0.0004
---------------------	--------	-----------------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance TIGHTENED SIM HEAD & RETESTED.

*DWB*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla D Platt*

SHAYLA D PLATT

Signature and Printed Name

02/02/2017  
Date