

Florida Department of Law Enforcement Alcohol Testing Program

FEB 10 2017

FDLE
Alcohol Testing Program

BCC

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 12:13

Date of Inspection: 01/25/2017

Serial Number: 80-000869
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:_____ Exp:_____ | 0.20g/210L Test (g/210L) Lot#:_____ Exp:_____ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____ |
|----------------------------|---|---|---|--|
| 0.000 | 0.051 / 0.051 | | | |
| 0.000 | 0.052 / 0.052 | | | |
| 0.000 | 0.052 / 0.052 | | | |
| 0.000 | 0.054 / 0.052 | | | |
| 0.000 | 0.054 / 0.053 | | | |
| 0.000 | 0.054 / 0.054 | | | |
| 0.000 | 0.054 / 0.056 | | | |
| 0.000 | 0.055 / 0.057 | | | |
| 0.000 | 0.056 / 0.055 | | | |
| 0.000 | 0.056 / 0.055 | | | |

| | | | |
|---------------------|-----------------|--|--|
| Standard Deviations | 0.0016 / 0.0020 | | |
|---------------------|-----------------|--|--|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance HIGH LEVEL CONDENSATION IN TU. UNKNOWN REASON Non-compliance:0.05G 210L WILL RECALIBRATE.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera

Signature and Printed Name

DAVID E REYES-RIVERA

01/25/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

FEB 10 2017

FDLE
Alcohol Testing Program

BOC

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 14:50

Date of Inspection: 01/25/2017

Serial Number: 80-000869
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:_____ Exp:_____ | 0.20g/210L Test (g/210L) Lot#:_____ Exp:_____ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____ |
|----------------------------|---|---|---|--|
| 0.000 / 0.000 | 0.049 | | | |
| 0.000 / 0.000 | 0.052 | | | |
| 0.000 / 0.000 | 0.052 | | | |
| 0.000 / 0.000 | 0.054 | | | |
| 0.000 / 0.000 | 0.055 | | | |
| 0.000 / 0.000 | 0.054 | | | |
| 0.000 / 0.000 | 0.054 | | | |
| 0.007 / 0.000 | 0.054 | | | |
| / 0.000 | 0.055 | | | |
| / 0.000 | 0.056 | | | |

| | | | |
|---------------------|--------|--|--|
| Standard Deviations | 0.0020 | | |
|---------------------|--------|--|--|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

00: Purge FailUNKNOWN REASON. 05: Control Outside ToleranceININSTRUMENT NEEDS REPAIRNon-compliance:INSTRUMENT NEEDS REPAIR.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes Rivera DAVID E REYES-RIVERA
Signature and Printed Name

01/25/2017
Date