

RECEIVED

Florida Department of Law Enforcement JUN 26 2017

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD
Time of Inspection: 08:29

Date of Inspection: 06/12/2017

Serial Number: 80-000874
Software: 8100.27

Table with 6 columns: Check or Test, YES, NO, Check or Test, YES, NO. Rows include Diagnostic Check (Pre-Inspection): OK, Minimum Sample Volume Check: OK, Alcohol Free Subject Test: 0.000, Interferent Detect Test: Interferent Detect.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Rows show test results for multiple samples.

Table with 5 columns: Standard Deviations, 0.0005, 0.0004, 0.0007, 0.0003.

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature: David E Reyes Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

06/12/2017
Date