

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

APR 04 2017

FDLE  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Vee

Agency: TITUSVILLE P.D.

Serial Number: 80-000996

Time of Inspection: 11:46

Date of Inspection: 04/04/2017

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.048	0.078 / 0.078	0.196	0.079
0.000	0.049	0.078 / 0.078	0.197	0.079
0.000	0.049	0.078 / 0.078	0.197	0.079
0.000	0.049	0.078 / 0.078	0.197	0.079
0.000	0.049	0.078 / 0.078	0.197	0.079
0.000	0.049	0.078 / 0.079	0.197	0.079
0.000	0.049	0.077 / 0.079	0.196	0.079
0.000	0.049	0.076 / 0.079	0.196	0.080
0.000	0.049	0.075 / 0.079	0.197	0.079
0.000	0.049	0.074 / 0.080	0.196	0.079

Standard Deviations	0.0003	0.0014 / 0.0006	0.0005	0.0003
---------------------	--------	-----------------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance - TIGHTENED SIM HEAD | CHECKED TUBING & RETESTED.

RDS BK

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

04/04/2017  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE P.D.  
Time of Inspection: 12:08

Date of Inspection: 02/02/2017

Serial Number: 80-000996  
Software: 8100.27

BK

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.049	0.079	0.194	0.079
0.000	0.049	0.079	0.196	0.079
0.000	0.048	0.079	0.196	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.195	0.079
0.000	0.048	0.078	0.195	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.195	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.195	0.079

Standard Deviations	0.0004	0.0004	0.0006	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

**Remarks:**

A F / M A: DID NOT PROVIDE ADEQUATE SAMPLE.

KMS

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

02/02/2017  
Date