

RECEIVED

Florida Department of Law Enforcement JUN 05 2017

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GULFPORT PD
Time of Inspection: 12:05

Date of Inspection: 06/05/2017

Serial Number: 80-001061
Software: 8100.27

Table with 6 columns: Check or Test, YES, NO, Check or Test, YES, NO. Rows include Diagnostic Check, Minimum Sample Volume Check, Alcohol Free Subject Test, and Interferent Detect Test.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Rows show test results for multiple samples.

Table with 5 columns: Standard Deviations, 0.0005, 0.0004, 0.0006, 0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

Handwritten signature in blue ink.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature of Danielle M Bell.

DANIELLE M BELL

Signature and Printed Name

06/05/2017
Date