

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUN 13 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: TARPON SPRINGS PD
Time of Inspection: 11:16

Date of Inspection: 06/13/2017

Serial Number: 80-001065
Software: 8100.27

bc

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG612405 Exp: 05/03/2018
0.000	0.000 / 0.050	0.081	0.200	0.080
0.000	0.000 / 0.050	0.081	0.200	0.080
0.000	0.000 / 0.050	0.081	0.201	0.079
0.000	0.000 / 0.050	0.080	0.201	0.079
0.000	0.000 / 0.050	0.081	0.200	0.080
0.000	0.000 / 0.050	0.081	0.200	0.080
0.000	0.000 / 0.049	0.081	0.200	0.080
0.000	0.000 / 0.050	0.080	0.200	0.080
0.000	0.000 / 0.050	0.080	0.201	0.079
0.000	0.000 / 0.050	0.079	0.201	0.080

Standard Deviations	0.0000 / 0.0003	0.0007	0.0005	0.0004
---------------------	-----------------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance CAP LEFT OFF AFTER TEMP CHECK.

Replaced cap & retested. ans

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

06/13/2017
Date

JD