

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JUN 08 2017

FDLE  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO  
Time of Inspection: 10:18

Date of Inspection: 06/08/2017

Serial Number: 80-001083  
Software: 8100.27

*VLC*

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.049	0.078	0.197 / 0.193	0.079
0.000	0.049	0.079	0.197 / 0.196	0.079
0.000	0.049	0.080	/ 0.195	0.079
0.000	0.049	0.079	/ 0.196	0.079
0.000	0.050	0.079	/ 0.195	0.079
0.000	0.049	0.080	/ 0.196	0.079
0.000	0.050	0.080	/ 0.195	0.079
0.000	0.049	0.079	/ 0.195	0.079
0.000	0.049	0.080	/ 0.196	0.079
0.000	0.050	0.079	/ 0.195	0.079

Standard Deviations	0.0004	0.0006	/ 0.0009	0.0000
---------------------	--------	--------	----------	--------

*JSM*

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:  
20: Purge Fail. INVESTIGATED REASON NO ISSUE FOUND

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

06/08/2017  
Date

*JSD*