

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 12 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

VCC

Agency: FL HIGHWAY PATROL

Serial Number: 80-001118

Time of Inspection: 12:27

Date of Inspection: 01/10/2017

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000 / 0.000	0.048	0.078 / 0.080		
0.000 / 0.000	0.048	0.079 / INT		
0.000 / 0.000	0.049	0.079		
INT / 0.000	0.049	INT		
/ 0.000	0.049			
/ 0.000	0.049			
/ 0.000	0.049			
/ 0.000	0.049			
/ 0.000	0.049			
/ 0.000	0.049			

Standard Deviations	0.0004			
---------------------	--------	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

00: Interferent Detect. 08: Interferent Detect, Interferen Detect. REASON UNKNOWN Non-compliance: INTERFERENT DETECT.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

BK

ADB

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

01/10/2017

Date

Florida Department of Law Enforcement Alcohol Testing Program

JAN 12 2017

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: *Ocala FHP*

SERIAL NUMBER: *80-001118*

DATE OF INSPECTION: *1/10/2017*

SOFTWARE: *8100-27*

Diagnostic Check (Pre-Inspection): OK	Yes	Date and/or Time Adjusted			No
Minimum Sample Volume Check: OK	Yes	Barometric Pressure Sensor Check: OK		Yes	
Alcohol Free Subject Test: 0.000	Yes	Mouth Alcohol Test: Slope Not Met		Yes	
Interferent Detect Test: Interferent Detect	Yes	Diagnostic Check (Post-Inspection): OK			No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000 / 0.000	0.048	0.078 / 0.080		
0.000 / 0.000	0.048	0.079 / INT		
0.000 / 0.000	0.049	0.079		
INT / 0.000	0.049	INT		
/ 0.000	0.049			
/ 0.000	0.049			
/ 0.000	0.049			
/ 0.000	0.049			
/ 0.000	0.049			
/ 0.000	0.049			

Standard Deviations	0.0004			
---------------------	--------	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks: 00: Interferent Detect. 08: Interferent Detect, Interferen Detect. REASON UNKNOWN Non-compliance: INTERFERENT DETECT.

BK *AMS*

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy _____ PATRICK J MURPHY
Signature and Printed Name

01/10/2017
Date