

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY S.O.  
Time of Inspection: 13:12

Date of Inspection: 11/30/2017

Serial Number: 80-001207  
Software: 8100.27

| Check or Test                               | YES | NO | Check or Test                          | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK       | Yes |    | Date and/or Time Adjusted              |     | No |
| Minimum Sample Volume Check: OK             | Yes |    | Barometric Pressure Sensor Check: OK   | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    | Mouth Alcohol Test: Slope Not Met      | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    | Diagnostic Check (Post-Inspection): OK | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201707D<br>Exp: 07/25/2019 | 0.08g/210L Test (g/210L)<br>Lot#:201707E<br>Exp: 07/25/2019 | 0.20g/210L Test (g/210L)<br>Lot#:201707C<br>Exp: 07/24/2019 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:AG626604<br>Exp: 09/22/2018 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.080   | 0.195   | 0.080   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.081   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.080   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.079   |
| 0.000                      | 0.049   | 0.081   | 0.197   | 0.080   |

|                     |        |        |        |        |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0003 | 0.0003 | 0.0006 | 0.0004 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

*DP*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

11/30/2017  
Date

*11/30/17  
JD*

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JUN 01 2017

FDLE  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY S.O.  
Time of Inspection: 12:34

Date of Inspection: 06/01/2017

Serial Number: 80-001207  
Software: 8100.27

*BC*

| Check or Test                               | YES | NO | Check or Test                          | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK       | Yes |    | Date and/or Time Adjusted              |     | No |
| Minimum Sample Volume Check: OK             | Yes |    | Barometric Pressure Sensor Check: OK   | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    | Mouth Alcohol Test: Slope Not Met      | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    | Diagnostic Check (Post-Inspection): OK | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201603D<br>Exp: 03/08/2018 | 0.08g/210L Test (g/210L)<br>Lot#:201611B<br>Exp: 11/15/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201604C<br>Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:AG626604<br>Exp: 09/22/2018 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.079   | 0.196   | 0.080   |
| 0.000                      | 0.048   | 0.080   | 0.198   | 0.079   |
| 0.000                      | 0.048   | 0.080   | 0.198   | 0.079   |
| 0.000                      | 0.048   | 0.081   | 0.198   | 0.079   |
| 0.000                      | 0.048   | 0.080   | 0.197   | 0.079   |
| 0.000                      | 0.048   | 0.080   | 0.195   | 0.079   |
| 0.000                      | 0.048   | 0.080   | 0.197   | 0.078   |
| 0.000                      | 0.049   | 0.080   | 0.197   | 0.078   |
| 0.000                      | 0.048   | 0.080   | 0.197   | 0.078   |
| 0.000                      | 0.049   | 0.080   | 0.197   | 0.079   |

|                     |        |        |        |        |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0004 | 0.0009 | 0.0006 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

*RMB*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

06/01/2017  
Date

*JD*