

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WEST PALM BEACH PD  
Time of Inspection: 13:29

Date of Inspection: 11/30/2017

Serial Number: 80-001235  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.050	0.082	0.198	0.079
0.000	0.050	0.082	0.199	0.079
0.000	0.050	0.082	0.199	0.079
0.000	0.050	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.050	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.199	0.080

Standard Deviations	0.0005	0.0000	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

*JA*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla D Platt*

SHAYLA D PLATT

Signature and Printed Name

11/30/2017  
Date

*11/30/17  
JA*

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

JUL 10 2017

**FDLE**

Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WEST PALM BEACH PD  
Time of Inspection: 12:54

Date of Inspection: 07/06/2017

Serial Number: 80-001235  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.048	0.078	0.202	0.082
0.000	0.048	0.081	0.202	0.081
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.049	0.080	0.199	0.081
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.080	0.200	0.081

Standard Deviations	0.0007	0.0007	0.0009	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

*POA*

*JD*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla D Platt*

SHAYLA D PLATT

Signature and Printed Name

07/06/2017  
Date