

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: EDGEWATER P.D.  
Time of Inspection: 11:43

Date of Inspection: 09/25/2017

Serial Number: 80-001243  
Software: 8100.27

| Check or Test                               | YES | NO | Check or Test                          | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK       | Yes |    | Date and/or Time Adjusted              |     | No |
| Minimum Sample Volume Check: OK             | Yes |    | Barometric Pressure Sensor Check: OK   | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    | Mouth Alcohol Test: Slope Not Met      | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    | Diagnostic Check (Post-Inspection): OK | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201707D<br>Exp: 07/25/2019 | 0.08g/210L Test (g/210L)<br>Lot#:201707E<br>Exp: 07/25/2019 | 0.20g/210L Test (g/210L)<br>Lot#:201707C<br>Exp: 07/24/2019 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:AG619605<br>Exp: 07/14/2018 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.049   | 0.082   | 0.199   | 0.080   |
| 0.000                      | 0.050   | 0.082   | 0.200   | 0.079   |
| 0.000                      | 0.049   | 0.082   | 0.199   | 0.081   |
| 0.000                      | 0.049   | 0.082   | 0.199   | 0.080   |
| 0.000                      | 0.049   | 0.081   | 0.199   | 0.081   |
| 0.000                      | 0.049   | 0.081   | 0.199   | 0.081   |
| 0.000                      | 0.049   | 0.082   | 0.199   | 0.080   |
| 0.000                      | 0.049   | 0.082   | 0.200   | 0.081   |
| 0.000                      | 0.050   | 0.082   | 0.199   | 0.081   |
| 0.000                      | 0.050   | 0.082   | 0.199   | 0.081   |

|                     |        |        |        |        |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0004 | 0.0004 | 0.0007 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

09/25/2017  
Date

*[Handwritten Signature]*

*9/25/17  
JL*

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 06 2017

FDLE *all*

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: EDGEWATER P.D.  
Time of Inspection: 12:54

Date of Inspection: 04/06/2017

Serial Number: 80-001243  
Software: 8100.27

*3K*

| Check or Test                               | YES | NO | Check or Test                          | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK       | Yes |    | Date and/or Time Adjusted              |     | No |
| Minimum Sample Volume Check: OK             | Yes |    | Barometric Pressure Sensor Check: OK   | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    | Mouth Alcohol Test: Slope Not Met      | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    | Diagnostic Check (Post-Inspection): OK | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201603D<br>Exp: 03/08/2018 | 0.08g/210L Test (g/210L)<br>Lot#:201611B<br>Exp: 11/15/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201604C<br>Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:AG626604<br>Exp: 09/22/2018 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.050   | 0.080   | 0.199   | 0.079   |
| 0.000                      | 0.050   | 0.081   | 0.199   | 0.080   |
| 0.000                      | 0.050   | 0.080   | 0.199   | 0.080   |
| 0.000                      | 0.049   | 0.080   | 0.200   | 0.079   |
| 0.000                      | 0.050   | 0.080   | 0.199   | 0.079   |
| 0.000                      | 0.049   | 0.080   | 0.199   | 0.079   |
| 0.000                      | 0.050   | 0.080   | 0.199   | 0.080   |
| 0.000                      | 0.050   | 0.080   | 0.199   | 0.080   |
| 0.000                      | 0.050   | 0.080   | 0.198   | 0.079   |
| 0.000                      | 0.050   | 0.080   | 0.199   | 0.079   |

|                     |        |        |        |        |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0003 | 0.0004 | 0.0005 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

*[Handwritten Signature]*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Patrick J Murphy*

PATRICK J MURPHY

Signature and Printed Name

04/06/2017  
Date