Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
JUN 13 2017
FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 11:01 Date of Inspection: 01/26/2017
Serial Number: 80-001248
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG605301 Exp: 02/22/2018
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.080	0.198	0.081
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.081	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.049	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.081
0.000	0.050	0.081	0.199	0.080
0.000	0.050	0.081	0.199	0.080

Standard Deviations	0.0005	0.0006	0.0006	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera
Signature and Printed Name

DAVID E REYES-RIVERA

01/26/2017
Date

RECEIVED
JUN 13 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 13:52 Date of Inspection: 01/24/2017
Serial Number: 80-001248
Software: 8100.27.10

VBC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK		No	Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: ____

RGM

Remarks:
FORMS LOAD TEST

Instrument erroneously down loaded software version on 80-001248. Version 8100.27.10 was loaded into Cobra.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera
DAVID E REYES-RIVERA
Signature and Printed Name

01/24/2017
Date

DR

FDLE/ATP Form 41 –Revised August 2005

RECEIVED
JUN 13 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 13:25
Date of Inspection: 01/24/2017
Serial Number: 80-001248
Software: 8100.27.10

BCC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK		No	Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: ____

PDM

Remarks:
TEST OF FORMS

Upon carefully insuring that software version 8100.27.00 was selected from the download menu the instrument continues to load 80-001248 with the 8100.27.10 software

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Reyes Rivera
Signature and Printed Name

DAVID E REYES-RIVERA

01/24/2017
Date

JD

FDLE/ATP Form 41 –Revised August 2005

Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
JUN 13 2017
FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Serial Number: 80-001248
Time of Inspection:10:49 Date of Inspection:01/24/2017 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG605301 Exp: 02/22/2018
0.000	0.049	0.080	0.199	0.081
0.000	0.049	0.080	0.200	0.081
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.199	0.081
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.081
0.000	0.051	0.080	0.200	0.081
0.000	0.050	0.081	0.200	0.081
0.000	0.050	0.080	0.199	0.081
Standard Deviations	0.0005	0.0003	0.0004	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks: form did not print after the inspection, had to retrieve it from Cobra

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Eduardo Reyes Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

01/24/2017

Date

FDLE/ATP Form 41 -- Revised August 2005

Reprinted From Database