

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 01 2017

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 17:27

Date of Inspection: 05/22/2017

Serial Number: 80-001250
Software: 8100.27

✓ BCC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.049	0.080	0.211 / 0.194	0.079
0.000	0.049	0.080	0.202 / 0.196	0.079
0.000	0.049	0.080	0.199 / 0.196	0.079
0.000	0.049	0.080	0.197 / 0.195	0.079
0.000	0.049	0.080	0.197 / 0.196	0.079
0.000	0.049	0.080	0.198 / 0.196	0.079
0.000	0.049	0.080	0.197 / 0.197	0.079
0.000	0.049	0.080	0.198 / 0.197	0.079
0.000	0.049	0.080	0.197 / 0.197	0.079
0.000	0.049	0.081	0.197 / 0.197	0.079

Standard Deviations	0.0000	0.0003	0.0043 / 0.0009	0.0000
---------------------	--------	--------	-----------------	--------

Plm

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance CONDENSATION IN THE TUBBING .

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

05/22/2017
Date

JD