

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUL 10 2017

FDLE

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: ORANGE COUNTY S.O.
Time of Inspection: 09:23

Date of Inspection: 07/06/2017

Serial Number: 80-001255
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.200 | 0.080 |
| 0.000 | 0.050 | 0.080 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.201 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.081 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.201 | 0.079 |
| 0.000 | 0.050 | 0.081 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.081 | 0.201 | 0.079 |
| 0.000 | 0.050 | 0.081 | 0.201 | 0.079 |
| 0.000 | 0.050 | 0.081 | 0.201 | 0.079 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0003 | 0.0006 | 0.0005 | 0.0003 |
|---------------------|--------|--------|--------|--------|


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

SP JD

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

07/06/2017
 Date