

Florida Department of Law Enforcement RECEIVED Alcohol Testing Program

MAR 02 2017 *YBC*

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 FDLE Alcohol Testing Program

Agency: ROCKLEDGE P.D. Serial Number: 80-001261
 Time of Inspection: 13:34 Date of Inspection: 03/02/2017 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG626604 Exp: 09/22/2018
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.080	0.199	0.079
0.000	0.049	0.081	0.199	0.079
0.000	0.049	0.081	0.199	0.079
0.000	0.049	0.081	0.198	0.079
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.080	0.197	0.078
0.000	0.049	0.081	0.198	0.079
0.000	0.049	0.081	0.198	0.079
0.000	0.049	0.081	0.198	0.079
0.000	0.049	0.081	0.198	0.079
Standard Deviations	0.0000	0.0009	0.0008	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

SHS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt SHAYLA D PLATT

 Signature and Printed Name

03/02/2017
Date