

RECEIVED

MAY 23 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 14:38

Date of Inspection: 05/17/2017

Serial Number: 80-001273
Software: 8100.27

✓

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.078 | 0.194 | 0.081 |
| 0.000 | 0.050 | 0.079 | 0.195 | 0.081 |
| 0.000 | 0.050 | 0.079 | 0.195 | 0.081 |
| 0.000 | 0.050 | 0.079 | 0.195 | 0.081 |
| 0.000 | 0.050 | 0.080 | 0.195 | 0.081 |
| 0.000 | 0.050 | 0.080 | 0.196 | 0.081 |
| 0.000 | 0.050 | 0.080 | 0.195 | 0.081 |
| 0.000 | 0.050 | 0.079 | 0.195 | 0.081 |
| 0.000 | 0.051 | 0.080 | 0.195 | 0.081 |
| 0.000 | 0.050 | 0.079 | 0.196 | 0.081 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0006 | 0.0005 | 0.0000 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

*SP
JD*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

05/17/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 14 2017 *✓*

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BK

Agency: PINELLAS COUNTY SO
Time of Inspection: 15:00

Date of Inspection: 02/14/2017

Serial Number: 80-001273
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.078 | 0.196 | 0.082 |
| 0.000 | 0.051 | 0.079 | 0.196 | 0.082 |
| 0.000 | 0.050 | 0.080 | 0.194 | 0.083 |
| 0.000 | 0.050 | 0.080 | 0.196 | 0.083 |
| 0.000 | 0.050 | 0.080 | 0.196 | 0.082 |
| 0.000 | 0.050 | 0.080 | 0.196 | 0.083 |
| 0.000 | 0.051 | 0.080 | 0.196 | 0.082 |
| 0.000 | 0.051 | 0.080 | 0.196 | 0.082 |
| 0.000 | 0.051 | 0.080 | 0.196 | 0.082 |
| 0.000 | 0.051 | 0.080 | 0.196 | 0.082 |
| 0.000 | 0.050 | 0.080 | 0.196 | 0.083 |
| Standard Deviations | 0.0005 | 0.0006 | 0.0006 | 0.0005 |

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:
A F / M A: TOO MUCH MOUTH ALCOHOL.

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

02/14/2017
Date