

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

AUG 29 2017

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Indian River Sheriff's Office
Time of Inspection: 06:00

Date of Inspection: 08/23/2017

Serial Number: 80-001328
Software: 8100.27

✓BC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		X	Date and/or Time Adjusted		X
Minimum Sample Volume Check: OK		X	Barometric Pressure Sensor Check: OK		X
Alcohol Free Subject Test: 0.000		X	Mouth Alcohol Test: Slope Not Met		X
Interferent Detect Test: Interferent Detect		X	Diagnostic Check (Post-Inspection): OK		X

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

Instrument does not turn on. Instrument will be sent to repair facility for evaluation.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.
I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E. Reyes-Rivera
Signature and Printed Name

8/23/2017
Date

FDLE/ATP Form 41 - Revised August 2005 Ref. 11D-8.004