

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

APR 03 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 11:10

Date of Inspection: 03/30/2017

Serial Number: 80-001344
Software: 8100.27

131K

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.051 | 0.081 | 0.199 | 0.081 |
| 0.000 | 0.051 | 0.081 | 0.199 | 0.081 |
| 0.000 | 0.052 | 0.081 | 0.200 | 0.080 |
| 0.000 | 0.052 | 0.082 | 0.199 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.200 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.199 | 0.080 |
| 0.000 | 0.052 | 0.081 | 0.200 | 0.080 |
| 0.000 | 0.052 | 0.081 | 0.200 | 0.080 |
| 0.000 | 0.052 | 0.082 | 0.200 | 0.081 |
| 0.000 | 0.052 | 0.081 | 0.200 | 0.080 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0005 | 0.0004 | 0.0005 | 0.0004 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

MS

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

03/30/2017
Date