

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Alachua PD
Time of Inspection:

Date of Inspection: 11/20/2017

Serial Number: 80-001720
Software: 8100.27

✓BLC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks: Instrument could not be inspected due to fault det. On optical bench cal. Was sent to CMI for repair.

The above instrument complies () does not comply (x) with Chapter 11D-8, FAC.
I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

 11/20/17
 Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ALACHUA PD
Time of Inspection: 08:47

Date of Inspection: 11/02/2017

Serial Number: 80-001720
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048	0.086 / 0.082		
0.000	0.047	0.089 / 0.090		
0.000	0.050	0.085 / 0.084		
0.000	0.051	0.084 / 0.083		
0.000	0.047	0.085 / 0.087		
0.000	0.048	0.087 / 0.083		
0.000	0.048	0.090 / 0.083		
0.000	0.053	0.089 / 0.082		
0.000	0.049	0.087 / 0.085		
0.000	0.050	0.081 / 0.087		
Standard Deviations	0.0019	0.0027 / 0.0026		

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance SIM CHECKED REPEAT, Control . SEE WRITTEN MEMONon-compliance:DID NOT PASS .08 WET.

Non-compliance - will Recheck Simulator on other instruments to determine cause of fluctuations. If not simulator dependent, will Cal Adjust + reinspect.



The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

11/02/2017
Date