

RECEIVED

AUG 31 2017

FDLE

Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD
Time of Inspection: 10:51

Date of Inspection: 08/31/2017

Serial Number: 80-002097
Software: 8100.27

BLC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG619605 Exp: 07/14/2018
0.000	0.049	0.081	0.202	0.079
0.000	0.049	0.081	0.202	0.078
0.000	0.049	0.081	0.203	0.079
0.000	0.049	0.081	0.203	0.079
0.000	0.049	0.081	0.203	0.079
0.000	0.049	0.082	0.202	0.079
0.000	0.049	0.081	0.203	0.079
0.000	0.049	0.081	0.202	0.079
0.000	0.049	0.081	0.203	0.079
0.000	0.050	0.081	0.202	0.079

Standard Deviations	0.0003	0.0003	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

Jan

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN
Signature and Printed Name

08/31/2017
Date

*8/31/17
JL*