

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

MAY 04 2017

FDLE

Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO

Serial Number: 80-003388

Time of Inspection: 14:32

Date of Inspection: 05/04/2017

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG626605 Exp: 09/22/2018
0.000	0.049	0.081	0.197	0.081
0.000	0.050	0.082	0.200	0.080
0.000	0.050	0.082	0.201	0.080
0.000	0.050	0.081	0.200	0.080
0.000	0.050	0.081	0.200	0.080
0.000	0.050	0.081	0.200	0.081
0.000	0.050	0.081	0.200	0.079
0.000	0.051	0.082	0.200	0.080
0.000	0.051	0.083	0.200	0.081
0.000	0.050	0.081	0.199	0.080

Standard Deviations	0.0005	0.0007	0.0010	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007      Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (    ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIELLE M BELL

Signature and Printed Name

05/04/2017

Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

MAR 08 2017

FDLE

Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

*BCC*

Agency: HILLSBOROUGH CO SO  
Time of Inspection: 20:16

Date of Inspection: 03/03/2017

Serial Number: 80-003388  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.048	0.080	0.208	/ 0.084
0.000	0.049	0.081	0.203	/ 0.085
0.000	0.049	0.080	0.202	/ 0.088
0.000	0.050	0.081	0.201	/ 0.088
0.000	0.050	0.081	0.202	/ 0.087
0.000	0.049	0.081	0.201	/ 0.087
0.000	0.049	0.081	0.201	/ 0.086
0.000	0.050	0.082	0.201	/ 0.086
0.000	0.049	0.081	0.202	/ 0.086
0.000	0.049	0.082	0.200	/ 0.086
Standard Deviations	0.0006	0.0006	0.0022	/ 0.0012

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0011 Number of Simulators Used: 5

**Remarks:**

08: Ambient Fail CLEARED AREA, Control Outside ToleranceUNK. UNKNOWN CONTROL OUTSIDE TOLERANCE, WILL CONDUCT FURTHER TESNon-compliance: UNKNOWN COT, FURTHER TESTING.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

*SP BK*

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Danielle M Bell*

DANIELLE M BELL

Signature and Printed Name

03/03/2017  
Date