

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

AUG 04 2017

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTATULA PD

Time of Inspection: 17:04

Date of Inspection: 08/03/2017

Serial Number: 80-006235

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018
0.000	0.048	0.079	0.198 / 0.202	0.079
0.000	0.048	0.079	0.198 / 0.206	0.079
0.000	0.048	0.079	0.196 / 0.207	0.079
0.000	0.049	0.079	0.194 / 0.207	0.079
0.000	0.048	0.079	0.191 / 0.207	0.079
0.000	0.048	0.079	0.188 / 0.207	0.079
0.000	0.049	0.079	0.186 / 0.207	0.079
0.000	0.049	0.079	0.184 / 0.207	0.079
0.000	0.049	0.079	0.182 / 0.206	0.080
0.000	0.049	0.079	0.179 / 0.206	0.079

Standard Deviations	0.0005	0.0000	0.0068 / 0.0015	0.0003
---------------------	--------	--------	-----------------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance LOST POWER TO SIMULATOR

Turned power back on, waited for it to heat, then retested. *DBS*

DB

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

08/03/2017
Date

*8/4/17
DB*