

RECEIVED

Florida Department of Law Enforcement JUL 10 2017

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES BEACH PD
Time of Inspection: 13:17

Date of Inspection: 07/06/2017

Serial Number: 80-006692
Software: 8100.27

Table with 6 columns: Check or Test, YES, NO, Check or Test, YES, NO. Rows include Diagnostic Check (Pre-Inspection): OK, Minimum Sample Volume Check: OK, Alcohol Free Subject Test: 0.000, Interferent Detect Test: Interferent Detect, and Diagnostic Check (Post-Inspection): OK.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test (g/210L), 0.08g/210L Test (g/210L), 0.20g/210L Test (g/210L), 0.08 g/210L Dry Gas Std Test (g/210L). Rows show test results for multiple samples.

Table with 5 columns: Standard Deviations, 0.0000, 0.0000, 0.0005, 0.0005.

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature: Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

07/06/2017
Date