

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 17 2017

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 11:27

Date of Inspection: 10/05/2017

Serial Number: 80-006693
Software: 8100.27

vsc

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626604 Exp: 09/22/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.078 | 0.194 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.195 | 0.079 |
| 0.000 | 0.050 | 0.079 | 0.195 | 0.079 |
| 0.000 | 0.050 | 0.079 | 0.195 | 0.079 |
| 0.000 | 0.049 | 0.080 | 0.195 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.194 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.195 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.194 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.194 | 0.079 |
| 0.000 | 0.049 | 0.080 | 0.194 | 0.079 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0006 | 0.0005 | 0.0000 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

PLM

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes RIVERA

DAVID E REYES-RIVERA

Signature and Printed Name

10/05/2017
Date

10/17/17