

Florida Department of Law Enforcement Alcohol Testing Program

JUN 08 2017

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP L
Time of Inspection: 09:59

Date of Inspection: 04/11/2017

Serial Number: 80-006769
Software: 8100.27

✓BAC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted	Yes	
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
INT / 0.000	0.050	0.079	0.196	0.079
/ 0.000	0.051	0.080	0.198	0.079
/ 0.000	0.051	0.081	0.199	0.079
/ 0.000	0.052	0.081	0.198	0.079
/ 0.000	0.051	0.081	0.198	0.079
/ 0.000	0.052	0.081	0.198	0.078
/ 0.000	0.051	0.081	0.198	0.079
/ 0.000	0.051	0.081	0.199	0.078
/ 0.000	0.052	0.081	0.199	0.079
/ 0.000	0.051	0.082	0.199	0.079

Agm

Standard Deviations	0.0006	0.0007	0.0009	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

Time-Date changed. 00: Interferent Detect WRONG SIM MUMULATOR. SYNC TIME WITH OTHER INSTRUMEN
TS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera DAVID E REYES-RIVERA
Signature and Printed Name

04/11/2017
Date

JD