



Florida Department of Law Enforcement

DEPARTMENT INSPECTOR FIELD NOTES

Agency: Tallahassee Police Dept. Instrument Serial Number: 80-002989

Department Inspection, Agency Inspection/Agency Contact, Records Audit. Includes checkboxes for Registration, Agency Inspection Notification, and Records Audit items.

Comments: A.I. notified us of a failed inspection due to interferent detect problems (see email). Will re-mix and re-run inspection

Instrument/Area, Equipment, Supplies. Includes checkboxes for Clean/Dry, Proper Number of Simulators, Distilled/Deionized Water, etc.

Comments: (Empty section for additional notes)

Action, Comments. Includes checkboxes for Instrument Compliance, Evidentiary Use, and Agency Inspection.

Signature of Alcohol Testing Program Staff Member: Patrick Murphy

Date: 5/25/2017

Signature of Alcohol Testing Program Staff Member

Date

Original - FDLE

Copy - Agency