



80-000752  
 Stability checks  
 11/2/17

INTOXILYZER 8000  
 Instrument Initialization  
 08:43 11/02/2017

MARION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000752  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:57
Control Test	0.200	08:57
Air Blank	0.000	08:58
Control Test	0.201	08:59
Air Blank	0.000	08:59
Control Test	0.201	09:00
Air Blank	0.000	09:00
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

Operator's Signature

MARION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000752  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:02
Control Test	0.079	09:03
Air Blank	0.000	09:04
Control Test	0.080	09:04
Air Blank	0.000	09:05
Control Test	0.079	09:05
Air Blank	0.000	09:06
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature

MARION COUNTY SO  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000752  
 11/02/2017  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:07
Control Test	0.049	09:08
Air Blank	0.000	09:08
Control Test	0.049	09:09
Air Blank	0.000	09:10
Control Test	0.049	09:10
Air Blank	0.000	09:11
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DS

Operator's Signature

Operator's Signature

Operator's Signature

11/3/17  
 DS

80-000752  
Flow Cal

Diagnostic Test

11/2/17

- Inadvertently ran Diagnostic test  
while trying to navigate back  
to diagnostic menu screen.

MARION COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000752  
11/02/2017  
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test OK  
RAM Test OK  
EEPROM Checksum Test OK  
Real Time Clock Test OK  
DSP Test OK  
Analytical Stability Test OK  
Internal Printer Test OK  
Modem Test OK  
Temperature Regulation Test OK

INTOXILYZER 8000  
Instrument Initialization  
06:35 11/02/2017

MARION COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000752  
11/02/2017  
Software: 8100.27

Flow Rate Calibration\*\*\*\*\*

1: Rate (Liters/min) = 5  
SORT(Diff) = 5.477  
2: Rate (Liters/min) = 15  
SORT(Diff) = 10.438  
3: Rate (Liters/min) = 30  
SORT(Diff) = 20.758

Dependent Data Scale Factor = 100000 L/min  
Independent Data Scale Factor = 256  
Rounded Slope = 628  
Rounded Intercept = -299314  
Correlation = 0.99640

WFM

11/3/17  
JQ